



Brazilian Journal of OTORHINOLARYNGOLOGY

www.bjorl.org



REVIEW ARTICLE

Diagnostic validity of methods for assessment of swallowing sounds: a systematic review[☆]



Karina Veríssimo Meira Taveira^{a,*}, Rosane Sampaio Santos^b,
Bianca Lopes Cavalcante de Leão^c, José Stechman Neto^b, Leandro Pernambuco^d,
Letícia Korb da Silva^e, Graziela De Luca Canto^{f,g}, André Luís Porporatti^f

^a Universidade Federal do Rio Grande do Norte (UFRN), Departamento de Morfologia, Centro de Biociências, Natal, RN, Brazil

^b Universidade Tuiuti do Paraná (UTP), Programa de Pós-graduação em Distúrbios da Comunicação, Curitiba, PR, Brazil

^c Universidade Tuiuti do Paraná (UTP), Departamento de Odontologia, Curitiba, PR, Brazil

^d Universidade Federal da Paraíba (UFPB), Departamento de Fonoaudiologia, João Pessoa, PB, Brazil

^e Instituto de Educação Luterana de Santa Catarina, Departamento de Fonoaudiologia, Joinville, SC, Brazil

^f Universidade Federal de Santa Catarina (UFSC), Departamento de Odontologia, Brazilian Centre for Evidence-based Research, Florianópolis, SC, Brazil

^g University of Alberta, Faculty of Medicine and Dentistry, School of Dentistry, Alberta, Canada

Received 26 September 2017; accepted 27 December 2017

Available online 3 February 2018

KEYWORDS

Deglutition;
Deglutition disorders;
Diagnosis;
Review

Abstract

Introduction: Oropharyngeal dysphagia is a highly prevalent comorbidity in neurological patients and presents a serious health threat, which may lead to outcomes of aspiration pneumonia, ranging from hospitalization to death. This assessment proposes a non-invasive, acoustic-based method to differentiate between individuals with and without signals of penetration and aspiration.

Objective: This systematic review evaluated the diagnostic validity of different methods for assessment of swallowing sounds, when compared to videofluoroscopy swallowing study to detect oropharyngeal dysphagia.

Methods: Articles in which the primary objective was to evaluate the accuracy of swallowing sounds were searched in five electronic databases with no language or time limitations. Accuracy measurements described in the studies were transformed to construct receiver operating characteristic curves and forest plots with the aid of Review Manager v. 5.2 (The Nordic Cochrane Centre, Copenhagen, Denmark). The methodology of the selected studies was evaluated using the Quality Assessment Tool for Diagnostic Accuracy Studies-2.

[☆] Please cite this article as: Taveira KV, Santos RS, Leão BL, Stechman Neto J, Pernambuco L, Silva LK, et al. Diagnostic validity of methods for assessment of swallowing sounds: a systematic review. Braz J Otorhinolaryngol. 2018;84:638–52.

* Corresponding author.

E-mails: karinnavm@hotmail.com, karinna@cb.ufrn.br (K.V. Taveira).

Peer Review under the responsibility of Associação Brasileira de Otorrinolaringologia e Cirurgia Cérvico-Facial.

PALAVRAS-CHAVE

Deglutição;
Distúrbios de
deglutição;
Diagnóstico;
Revisão

Results: The final electronic search revealed 554 records, however only 3 studies met the inclusion criteria. The accuracy values (area under the curve) were 0.94 for microphone, 0.80 for doppler, and 0.60 for stethoscope.

Conclusion: Based on limited evidence and low methodological quality because few studies were included, with a small sample size, from all index testes found for this systematic review, doppler showed excellent diagnostic accuracy for the discrimination of swallowing sounds, whereas microphone-reported good accuracy discrimination of swallowing sounds of dysphagic patients and stethoscope showed best screening test.

© 2018 Associação Brasileira de Otorrinolaringologia e Cirurgia Cérvico-Facial. Published by Elsevier Editora Ltda. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

Validade diagnóstica dos métodos de avaliação dos sons de deglutição: uma revisão sistemática

Resumo

Introdução: A disfagia orofaríngea é uma comorbidade altamente prevalente em pacientes neurológicos e representa uma séria ameaça à saúde, pode levar a desfechos como pneumonia por aspiração, hospitalização e até morte. A avaliação propõe um método não invasivo, acústico, para diferenciar entre indivíduos com e sem sinais de penetração e aspiração.

Objetivo: Esta revisão sistemática analisou a validade diagnóstica de diferentes métodos para avaliação dos sons de deglutição, quando comparados com a videofluoroscopia da deglutição para detectar disfagia orofaríngea.

Método: Artigos nos quais o objetivo principal era avaliar a acurácia dos sons de deglutição foram pesquisados em cinco bancos de dados eletrônicos sem limitações de idioma ou tempo de publicação. As medidas de acurácia descritas nos estudos foram transformadas para construir curvas ROC (*Receptor Operating Characteristic*) e gráfico em floresta (*forest plot*) com o auxílio do software *Review Manager v. 5.2* (The Nordic Cochrane Centre, Copenhagen, Dinamarca). A metodologia dos estudos selecionados foi avaliada com a ferramenta Avaliação da Qualidade de Estudos de Acurácia de Testes Diagnósticos-2.

Resultados: A busca eletrônica final resultou na identificação de 554 artigos; no entanto, apenas três estudos preencheram os critérios de inclusão. Os valores de acurácia (área abaixo da curva) foram 0,94 para microfone, 0,80 para doppler e 0,60 para estetoscópio.

Conclusão: Baseado nas evidências limitadas e da baixa qualidade metodológica, pois foram poucos os estudos incluídos, e com pequeno tamanho amostral, de todos os testes diagnósticos (*index testes*) encontrados para essa revisão sistemática o doppler mostrou excelente acurácia diagnóstica na discriminação dos sons de deglutição, o microfone demonstrou uma boa acurácia na discriminação dos sons de pacientes disfágicos e o estetoscópio revelou o melhor teste de triagem.

© 2018 Associação Brasileira de Otorrinolaringologia e Cirurgia Cérvico-Facial. Publicado por Elsevier Editora Ltda. Este é um artigo Open Access sob uma licença CC BY (<http://creativecommons.org/licenses/by/4.0/>).

Introduction

Swallowing is characterized by an intricate neuromuscular mechanism that requires a sequence of biomechanical activities, resulting in the passage of liquids and solids from mouth to stomach, avoiding the airway.^{1,2} Dysphagia may bring serious and potentially fatal health consequences, which negatively impact the well-being, safety, quality of life, and safety of patients.^{3,4} Aspiration is one of the most serious manifestations of oropharyngeal dysphagia, and may be the cause of undernourishment, chest infection, prolonged hospital stay and, lastly, mortality.⁵ Prevalence measurements for dysphagia diverge, depending upon the

etiology and patient's age, but estimates as high as 38% for lifetime prevalence have been reported in those over 65-years-old.⁶

To avoid unfavorable health results, detecting dysphagia early is crucial as well as to initiate an early referral for diagnosis and treatment to minimize health threats. The test named Videofluoroscopic Swallowing Study (VFSS), which consists of asking a patient to swallow different foods and liquids that contain a radiopaque contrast agent while observed by a trained professional is often considered the standard reference to determine if dysphagia exists.⁷⁻¹¹ For this test, kinematic X-ray data for physiological swallow impairment and subsequent misdirection of swallowed

material^{12,13} are observed by a trained examiner. However, frequent VFSS test repetitions are not recommended due to high radiation exposures.¹⁴

There is a noninvasive method that has been proposed by acoustic means for swallowing analysis. Microphones and/or accelerometers are used to record breath and swallowing sounds, which are examined using digital signal processing methods. Swallowing sounds have been widely associated with pharyngeal reverberations arising from opening and closing of valves (oropharyngeal, laryngeal and esophageal valves), action of numerous pumps (pharyngeal, esophageal, and respiratory pumps) and vibrations of the vocal tract.¹⁵

Literature on swallowing sounds to supplement the clinical evaluation of dysphagia has shown promising results.^{16,17} There are no studies correlating the diagnostic accuracy as a method for the detection of swallowing sounds. Based on the above, the aim of this systematic review was to answer the focused question: "What is the diagnostic validity of different methods for assessment of swallowing sounds, when compared to VFSS, for detecting oropharyngeal dysphagia?"

Methods

Protocol and registration

PRISMA statement¹⁸ was used to guide the execution of this systematic review; and the protocol was registered on International Prospective Register of Systematic Reviews (PROSPERO) database (Registration n° CRD42016052771).

Eligibility criteria

We have included diagnostic validity studies, which used different methods for assessment of swallowing sounds compared to the reference standard: videofluoroscopy (VFSS). Different methods for assessment of swallowing sounds could include ultrasound, acoustic analysis, cervical auscultation, swallowing accelerometers signals, and the Doppler effect. Previous studies from all languages and with no restrictions regarding age, sex and time of publication were included.

Exclusion criteria

Articles were excluded from review based on the following criteria: (1) Studies in animals; (2) Studies that did not perform ultrasound, acoustic analysis, cervical auscultation, swallowing accelerometers signals or Doppler effect; (3) Studies that did not compare methods of diagnosis of swallowing for both control and dysphagic group with the VFSS reference standard; (4) Studies that did not present validity measurements (sensitivity and specificity) or did not present data enough to calculate them; (5) Reviews, letters, conference, abstract, personal opinions.

Information sources

A computerized literature search was conducted in five main databases, such as Cochrane, Latin American and Caribbean Health Sciences (LILACS), PubMed (including Medline),

Scopus, Web of Science; and three grey literature databases (Google Scholar, OpenGrey, and ProQuest Dissertation and Thesis). More information on the search strategies is provided in [Appendix 1](#). Furthermore, the reference lists of the selected articles were inspected for additional literature. Relevant papers on this topic were also requested from experts in the field. The references were managed and the duplicates hits were removed with the aid of EndNote Basic X7® Software (Thompson Reuters, New York, NY, USA). We conducted all searches on October 8th, 2016. An updated search with the same word combinations for each database above mentioned was performed on January 25th, 2017.

Study selection

Two independent reviewers (K.V.M.T. and R.S.S.) made the first preselecting cut by screening all articles on title and abstract. Studies which did not appear to meet the eligibility criteria were excluded. Next, they independently screened full texts of this initial set of articles. Any disagreements were resolved through discussion or referral to a third author (B.L.C.L.).

Data collection process

Data extraction was performed by one author (K.V.M.T.) and checked by a second (R.S.S.). Disagreements were resolved through discussion. A third author (B.L.C.L.) became involved, when needed, to make a final decision.

Data items

The data collected consisted of study authors, year of publication, country, design, mean age and range, sample size, number of patients, number of observations, index test, reference test, description, outcomes, and conclusions. Efforts were made to contact the authors to recover any unpublished data, if the required data were not complete.

Risk of bias in individual studies

The included studies were assessed for methodological quality using the Quality Assessment Tool for Diagnostic Accuracy Studies (QUADAS-2).¹⁹ The following four methodological domains were measured for each trial: patient selection, index test, reference standard, flow of patients through the study, and timing of the tests.

Two independent reviewers (K.V.M.T. and R.S.S.) used its critical appraisal criteria to analyze all included articles, scoring each criterion with 'yes', 'no', or 'unclear'. Disagreements by discussion with a third author (B.L.C.L.) were made when necessary. Figures of the risk of bias assessment for all included studies were generated with Review Manager 5.3 (RevMan 5.3, The Nordic Cochrane Centre, Copenhagen, Denmark).

Summary measures

Sensitivity and specificity of the diagnostic tests were the main outcomes evaluated. Positive Predictive Value (PPV),

Negative Predictive Value (NPV), Positive Likelihood Ratio (LR+), Negative Likelihood Ratio (LR-), Diagnostic Odds Ratio (DOR), and Youden's index were secondary outcomes. The cutoff values used to interpret these data are presented in [Appendix 2](#).

Synthesis of results

Cochrane Collaboration guidelines²⁰ was used to combine individual results by means of a systematic review, with Restricted Maximum-Likelihood (REML) estimation and the DerSimonian pooled method. All statistical analysis was crude, without adjustment for potential confounders. Some of the required data were not specified in the articles, so we calculated them. Review Manager 5.3 (RevMan 5.3, The Nordic Cochrane Centre, Copenhagen, Denmark) was used to draw Receiver Operating Characteristic (ROC) curves, graphs, and forest plots. Heterogeneity within studies was evaluated either by considering clinical, methodological, and statistical characteristics or by using inconsistency Indexes (I^2), whereas a value greater than 50% was considered an indicator of substantial heterogeneity between studies, and a random effect applied. The significance level was set at 5%.²¹

Risk of bias across studies

Clinical, methodological, and statistical heterogeneity were explored among studies.

Results

Study selection

Systematic searches yielded 554 results, as shown in the PRISMA ([Fig. 1](#)). After removing the duplicates, a comprehensive evaluation of the 355 abstracts was performed and 330 articles were excluded, resulting in 25 articles for full-text reading. Grey literature search identified 253 studies, where none of the studies were selected. Also, after hand-search of the reference lists and articles provided by experts, no additional studies were included.

Therefore, 25 articles were retrieved for full-text reading. Twenty-two of them were excluded ([Appendix 3](#)). Finally, three studies remained and were included in the qualitative synthesis.

Study characteristic

The three included studies were published in 2004, 2013, and 2015. They were conducted in Brazil,²² Japan²³ and United Kingdom.²⁴ The sample size ranged from 10 to 30 healthy patients and 14 to 70 dysphagic patients. The index tests used were microphone²³ stethoscope with a microphone inserted into tubing at the bifurcation²⁴ and sonar Doppler.²²

The consistencies and viscosities of the material used to execute the reference test also varied. Abdulmassih et al.²² used three consistencies: liquid, 70 mL water and 30 mL of 100% barium sulfate; pudding, 70 mL of water, 30 mL of

barium sulfate; solids, club social biscuits soaked in barium during the reference test. Jayatilake et al.²³ used water swallow test to group healthy and 3 mL water mixed 25% barium group dysphagic during the reference test and Leslie et al.²⁴ used two consistencies, 3 boluses each of 5, 20 mL thin barium and 5 mL yogurt during the reference test. The liquid bolus volumes in the reference test varied from 3²⁴ to 70 mL.²² The size of the solid boluses was expressed in club social biscuits soaked in barium. Characteristics of included studies are described in [Table 1](#).

Risk of bias within studies

Although no studies fulfilled all criteria of risk of bias, the studies methods were very homogeneous and all possessed low risk of bias for applicability concerns ([Appendix 4](#)). For every study, item one of domain one that discuss risk of bias of patient selection was scored as high risk of bias, because each study recruited an experimental sample, without randomization of the enrolled patients. Item one of domain "index test" was scored "unclear" for two studies, because of results of screening or the interpretation of the test. The items reference test, flow and timing for the three included studies were scored "low". [Fig. 2](#) summarizes QUADAS-2 assessments.

Results of individual studies


Abdulmassih et al.²² evaluated acoustic analysis of swallow on 30 healthy patients and 30 dysphagic patients using a sonar Doppler compared to the VFSS. The analysis of variance of the averages found in each variable – frequency, intensity and duration of swallowing – shows there was a significant correlation when compared to the healthy individual curve.

Jayatilake et al.²³ evaluated real-time swallowing sound on 8 healthy subjects and 31 dysphagic patients using a microphone compared to the VFSS. 71 dry swallows the automatic swallow recognition algorithm achieved sensitivity 93.9%; algorithm automatically detected all or some of the swallowing events of all the 31 subjects dysphagic, and the overall detection accuracy for the 92 swallowing episodes was 79.3%.

Leslie et al.²⁴ evaluated acoustic analysis of swallow on 10 healthy subjects and 10 dysphagic patients using a microphone compared to the VFSS. When the assessors were asked whether the swallow was normal or abnormal, the sensitivity and specificity were low (sensitivity 62%, specificity 66%). When consensus was reached among the raters, the majority consensus gave 90% specificity, 80% sensitivity for detecting swallow normality.

Synthesis of results

All three articles^{22–24} contained enough data to be included in our systematic review. A diagnostic test validity table was constructed using the data extracted from each study ([Table 2](#)). In this table, all prevalence and accuracy measurements (sensitivity, specificity, PPV, NPV, LR+, LR-, DOR, and Youden's index) are presented. The total sample size for



¹Adapted from PRISMA.

Figure 1 Flow diagram of literature search and selection criteria.¹

this systematic review was 117 subjects, 48 healthy subjects and 69 dysphagic patients.

Sensitivity and specificity for different selected studies varied substantially. The diagnostic accuracy (sensitivity, specificity, and 95% Confidence Interval) of each study included in this systematic review is shown in Fig. 3. Sensitivity and specificity for microphone was 94% and 25% (95% CI 0.79–0.99) respectively,²³ sensitivity and specificity for Doppler was 80% and 100% (95% CI 0.61–0.92) respectively²² and sensitivity and specificity for stethoscope was 62% and 66% (95% CI 0.32–0.84) respectively.²⁴

The orders of the best diagnostic tests for dysphagic patients were microphone, Doppler and stethoscope. The

orders of the best diagnostic tests for healthy patients were Doppler, stethoscope and microphone.

Additional analysis

We chose to showcase the systematic review results in ROC curves (Fig. 4). Because of differences in the assessment of swallowing sounds methods, no cutoff point measures were justified and thus no threshold effect was possible; therefore, a symmetric curve was applied.

Regarding PPV values, the highest PPV values reported for microphone and Doppler^{22,23} showing that these

Table 1 Summary of descriptive characteristics and outcomes of interest of the included studies ($n = 3$).

| Author, year, country | Mean age range (years) | Sample size n° of patients | Sample size n° of observations | Index test | Reference test | Description | Outcome | Conclusion |
|--|------------------------|----------------------------|--------------------------------|------------|---|---|--|---|
| Abdulmassih et al., 2013, Brazil ¹⁸ | 46.4 (28–62) healthy | 30 healthy | 30 healthy | Doppler | VFSS Swallow material: liquid, 70 mL water and 30 mL of 100% barium sulfate; pudding, 70 mL of water, 30 mL of barium sulfate; solids, club social biscuits soaked in barium | Acoustic analysis of swallow | The prevalence in the dynamic evaluation of swallowing VFSS was by changes in the oral phase of swallowing. The analysis of variance of the averages found in each variable – frequency, intensity and duration of swallowing – shows there was a significant correlation when compared to the healthy individual curve. | In patients with SCA, the mean initial frequency, initial intensity, and final intensity were higher and the time and peak frequency were lower, demonstrating a pattern of cricopharyngeal opening very close to that found in normal populations. |
| | 44.9 (28–62) dysphagic | 30 dysphagic | 30 dysphagic | | | | | |
| Jayatilake et al., 2015, Japan ¹⁹ | (22–39) healthy | 15 healthy | 8 healthy | Microphone | VFSS Swallow material: group healthy, water swallow test; group dysphagic, 3 mL water mixed 25% barium | Real-time swallowing sound-processing algorithm for the automatic screening, quantitative evaluation, and the visualization of swallowing ability | 71 dry swallows the automatic swallow recognition algorithm achieved sensitivity 93.9% healthy subjects; algorithm automatically detected all or some of the swallowing events of all the 31 subjects dysphagic, and the overall detection accuracy for the 92 swallowing episodes was 79.3% | Swallowscope can analyze swallowing sounds in realtime and generate quantitative results: the number of swallows and the swallowing duration, which can assist bedside screening, and share them through a cloud-based system. We achieved very good performances in terms of both the positive predictive value and sensitivity. |
| | 68.8 dysphagic | 70 dysphagic | 31 dysphagic | | | | | |

Table 1 (Continued)

| Author, year, country | Mean age range (years) | Sample size n° of patients | Sample size n° of observations | Index test | Reference test | Description | Outcome | Conclusion |
|---|------------------------|----------------------------|--------------------------------|-------------|--|------------------------------|--|---|
| Leslie et al., 2004, United Kingdom ²⁰ | 72 (24–78) healthy | 10 healthy | 10 healthy | Stethoscope | VFSS Swallow material: 3 boluses each of 5, 20 mL thin barium and 5 mL yogurt | Acoustic analysis of swallow | Comparison with radiological defined aspiration/penetration yielded 66% specificity, 62% sensitivity, and majority consensus gave 90% specificity, 80% sensitivity for detecting normality of a swallow, when consensus is reached among the raters. | Improving the poor raters would improve the overall accuracy of this technique in predicting abnormality in swallowing. The group consensus correctly identified 17 of the 20 clips so we may speculate that the swallow sound contains audible cues that should in principle permit reliable classification. |
| | 78 (65–90) dysphagic | 14 dysphagic | 10 dysphagic | | | | | |

VFSS, Videofluoroscopic Swallowing Study; SCA, Spinocerebellar Ataxia; RSST, Repetitive Saliva Swallowing Test.




Figure 2 Results from QUADAS-2 study quality assessment (A, risk of bias graph; B, risk of bias summary).




Figure 3 Coupled forest plot of the sensitivity and specificity in videofluoroscopic swallowing studies compared and swallow sounds ($n = 3$).

techniques were able to discriminate swallowing sounds without lesion data 100% of the time. Doppler also reported to have the highest NPV, distinguishing control patients from those with acoustical analysis the swallowing 100% of the time.²²

Regarding LR values, 3 studies showed LR+ greater than 1.00 for swallowing sounds with stethoscope, microphone and Doppler,^{22–24} which means that all methods captured argue for dysphagia.²⁵ The highest LR+ value was reported for Doppler (LR+ = ∞)²² followed by stethoscope (LR+ = 1.85)²⁴ and microphone (LR+ = 1.32)²³ LR– values closer to 0 were reported for Doppler and microphone,^{22,23}

which means a low probability of disease when it is absent in the examination.²⁵

Finally, Doppler and stethoscope reported the highest DOR,^{22,24} indicating better discriminatory test performance.²⁶ The Doppler reported good Youden's Index (0.80).²²

Risk of bias across studies


The main methodological limitations across studies were related to poor reporting for Quadas-2 item "risk of bias of patient selection" scored with high risk of bias.

Table 2 Diagnostic test validity data ($n = 3$).

| Group | Author, year | Dysphasic sample size | Control sample size | Prevalence (%) | Sensitivity (%) | Specificity (%) | PPV | NPV | LR+ | LR– | DOR | Youden's index |
|------------------|--|-----------------------|---------------------|-------------------|-------------------|-------------------|-------------------|-------------------|-------------------|-------------------|-------------------|-------------------|
| Doppler/VFSS | Abdulmassih et al., 2013 ¹⁸ | 24 | 30 | 50.0 ^a | 80.0 ^a | 100 ^a | 1.00 ^a | 0.83 ^a | ∞^a | 0.20 ^a | ∞^a | 0.80 ^a |
| Microphone/VFSS | Jayatilake et al., 2015 ¹⁹ | 31 | 8 | 79.4 | 93.9 | 29.1 ^a | 0.83 | 0.55 ^a | 1.32 ^a | 0.20 ^a | 0.20 ^a | 0.23 ^a |
| Stethoscope/VFSS | Leslie et al., 2004 ²⁰ | 14 | 10 | 58.3 ^a | 62.0 | 66.0 | 0.71 ^a | 0.55 ^a | 1.82 ^a | 0.57 ^a | 3.16 ^a | 0.28 ^a |

VFSS, Videofluoroscopic Swallowing Study; PPV, Positive Predictive Value; NPV, Negative Predictive Value; LR+, Positive Likelihood Ratio; LR–, Negative Likelihood Ratio; ∞ , infinite.

^a Data calculated by the authors from information available in the article.

**Figure 4** ROC curves of the sensitivity and specificity in videofluoroscopic swallowing studies compared and swallow sounds.

Additionally, a poor agreement across the index test's observers was related at two studies, or it was unclear, resulting in a risk of bias of index test.

Discussion

This systematic review investigated different methods for assessment of swallowing sounds comparing VFSS among patients with oropharyngeal dysphagia. While several non-instrumented screening procedures have been adopted in medical centers worldwide, efforts to develop improved dysphagia screening methods with both high sensitivity and specificity are currently in development. In this systematic review the presented Doppler has good sensitivity and specificity to capture swallowing sounds and can be used as a method of diagnosis of dysphagic patients and healthy subjects, being a inexpensive and non-invasive method in relation to the reference standard VFSS. The presented microphone has high sensitivity and low specificity to capture swallowing sounds and can be used as a method of diagnosis of dysphagic patients, while the presented stethoscope has low sensitivity and low specificity to capture swallowing sounds and can be used as a method of screening of dysphagic patients.

VFSS is a radiologic procedure, whereby subjects ingest small amounts of barium-coated boluses while X-rays penetrate the subject and the resultant images are video-recorded. The VFSS test allows immediate visual inspection; however, it is time-consuming, non-portable and results in some radiation exposure.²⁷ Due to radiation exposure, the VFSS procedure is limited in duration and cannot be frequently repeated.²⁸ Thus, new techniques need to be developed to help assess the performance of the swallowing mechanism.

Some reproducible characteristic sound patterns have been reported to be heard during auscultation

of swallows with a stethoscope,²⁹ microphone^{30,31} or accelerometer^{30,32,33} and Doppler.^{22,34}

We found only 3 eligible studies with data concerning Doppler, microphone and stethoscope.^{22–24} No data were found for accelerometry. The sensitivity and specificity index values of tests varied among the included studies. The differences in these scores probably reflect the method of sounds of swallowing that are captured, placed on the neck to detect cervical sounds generated during the swallow and breath sounds pre- and post-swallow. Microphones and/or accelerometers are used to record breath and swallowing sounds, which are analyzed then using digital signal processing techniques. The research on using swallowing sounds to supplement the clinical evaluation of dysphagia has shown promising results.¹⁶

The PPV and NPV values confirmed the ROC curve results (Fig. 2). The higher PPV related to Doppler showed a lower risk of false positive results. In addition, a high NPV noticed in Doppler evaluations indicates that there is also a lower risk of under-diagnosis. It is important to emphasize that the prevalence of a disease can affect PPV and NPV values. When prevalence is high, true-positive results are more likely to be found in the population instead of false-positives, increasing the PPV and decreasing the NPV, respectively.³⁵ Similarly, the DOR values of index tests reported indicate that Doppler resulted in better discriminatory test performance²² and satisfied the criteria required for an excellent diagnostic test. Finally, LR+ and LR– values expressed better diagnostic accuracy for Doppler.¹⁷ The Doppler reported good Youden's Index (0.80).²²

To the best of the authors' knowledge, this is the first systematic review to validate sensitivity and specificity of sounds of swallowing. These values, added to PPV, NPV, LR+, LR–, ROC curve, and Youden's index analyses, were used for diagnostic accuracy.

In this study, the best diagnostic accuracy results were reported when using Doppler for captured the swallowing sound and can be used as a method of evaluation of dysphagic patients and healthy subjects, being a cheap and non-invasive method in relation to the reference standard VFSS.

Some methodological limitations of this review should be considered. First, different methods of catching swallowing sounds were used. Furthermore, 22 studies had to be excluded due to the lack of compared methods of diagnosis of swallowing for both control and dysphagic group with the reference test may be due to exposure to radiation to healthy patients. Finally, regarding the risk of bias from the included studies, no information about blinding was reported by most of the studies. Also, the preponderance of studies failed to report if the standard reference results were interpreted without knowledge of the results of the index test.

Studies that did not compare an index test with the reference test were not included, because only an acceptable reference test can prove the clinical relevance and reduce the risks of both false positive as well as the false-negative findings. Studies lacking comparisons of methods of diagnosis of swallowing for both control and dysphagic group with the reference test were also excluded.

Conclusion

Based on limited evidence and low methodological quality because few studies were included, with a small sample size, from all index testes found for this systematic review, the Doppler showed excellent diagnostic accuracy on the discrimination of swallowing sounds, whereas the microphone reported good sensitivity for discrimination of swallowing sounds of dysphagic patients and the stethoscope showed best screening test on the discrimination of swallowing sounds. Further studies with different methods for evaluation of swallowing sounds and with more representative samples are fully encouraged. Additional studies on this topic with a paired control group are also recommended.

Conflicts of interest

The authors declare no conflicts of interest.

Appendix 1. Database search strategy.

| Database | Search (October 8th 2016; updated on January 25th, 2017) |
|----------|---|
| Cochrane | "deglutition" OR "deglutitions" OR "swallowing" OR "swallowings" OR "swallows" OR "swallow" in Title, Abstract, Keywords and videofluoroscopy OR "Videofluoroscopy Swallowing Study" OR VFSS OR "videofluoroscopy study" OR "swallowing videofluoroscopy" OR videofluoroscopic OR "videofluoroscopic swallowing" OR "videofluoroscopic swallowing study" OR fluoroscopy OR fluoroscopies in Title, Abstract, Keywords and ultrasonography OR ultrasound OR ultrasonics OR "duplex doppler ultrasonography" OR "doppler duplex ultrasonography" OR "acoustic analysis" OR "acoustical analysis" OR auscultation OR "cervical auscultation" OR accelerometry OR "swallowing accelerometry signals" OR "swallowing sounds" OR "swallow sounds" OR "signal processing" OR acoustics OR acoustic OR "doppler effect" OR "doppler shift" OR "sonar doppler" OR microphone in Title, Abstract, Keywords in Trials |
| LILACS | (tw:("deglutition" OR "deglutitions" OR "swallowing" OR "swallowings" OR "swallows" OR "swallow")) AND (tw:(videofluoroscopy OR "videofluoroscopy swallowing study" OR vfss OR "videofluoroscopy study" OR "swallowing videofluoroscopy" OR videofluoroscopic OR "videofluoroscopic swallowing" OR "videofluoroscopic swallowing study" OR fluoroscopy OR fluoroscopies)) AND (tw:(ultrasonography OR ultrasound OR ultrasonics OR "duplex doppler ultrasonography" OR "doppler duplex ultrasonography" OR "acoustic analysis" OR "acoustical analysis" OR auscultation OR |

| | | | |
|--------|---|----------------|--|
| PubMed | <p>"cervical auscultation" OR accelerometry OR "swallowing accelerometry signals" OR "swallowing sounds" OR "swallow sounds" OR "signal processing" OR acoustics OR acoustic OR "doppler effect" OR "doppler shift" OR "sonar doppler" OR microphone)) ("deglutition"[MeSH Terms] OR "deglutition"[All Fields] OR "deglutitions"[All Fields] OR "swallowing"[All Fields] OR "swallowings"[All Fields] OR "swallows"[MeSH Terms] OR "swallows"[All Fields] OR "swallow"[All Fields]) AND (videofluoroscopy[All Fields] OR "videofluoroscopy swallowing study"[All Fields] OR VFSS OR "videofluoroscopy study"[All Fields] OR "swallowing videofluoroscopy"[All Fields] OR videofluoroscopic[All Fields] OR "videofluoroscopic swallowing"[All Fields] OR "videofluoroscopic swallowing study"[All Fields] OR fluoroscopy[MeSH Terms] OR fluoroscopy[All Fields] OR fluoroscopies) AND (ultrasonography[Subheading] OR ultrasonography[All Fields] OR ultrasound[All Fields] OR ultrasonography[MeSH Terms] OR ultrasound[All Fields] OR ultrasonics[MeSH Terms] OR ultrasonics[Text Word] OR ultrasonics[All Fields] OR "duplex doppler ultrasonography"[All Fields] OR "doppler duplex ultrasonography"[All Fields] OR "acoustic analysis"[All Fields] OR "acoustical analysis"[All Fields] OR auscultation[MeSH Terms] OR auscultation[All Fields] OR "cervical auscultation"[All Fields] OR accelerometry[MeSH Terms] OR accelerometry[All Fields] OR "swallowing accelerometry signals"[All Fields] OR "swallowing sounds"[All Fields] OR "swallow sounds"[All Fields] OR "signal processing"[All Fields] OR acoustics[MeSH Terms] OR acoustics[All Fields] OR acoustic[All Fields] OR "doppler effect"[MeSH Terms] OR doppler effect[Text Word] OR "doppler effect"[All Fields] OR "doppler shift" OR "shift, doppler" OR "effect, doppler" OR "sonar doppler"[All Fields] OR microphone[All Fields])</p> | Scopus | <p>(TITLE-ABS-KEY("deglutition" OR "deglutitions" OR "swallowing" OR "swallowings" OR "swallows" OR "swallow") AND TITLE-ABS-KEY(videofluoroscopy OR "Videofluoroscopy Swallowing Study" OR VFSS OR "videofluoroscopy study" OR "swallowing videofluoroscopy" OR videofluoroscopic OR "videofluoroscopic swallowing" OR "videofluoroscopic swallowing study" OR fluoroscopy OR fluoroscopies) AND TITLE-ABS-KEY(ultrasonography OR ultrasound OR ultrasonics OR "duplex doppler ultrasonography" OR "doppler duplex ultrasonography" OR "acoustic analysis" OR "acoustical analysis" OR auscultation OR "cervical auscultation" OR accelerometry OR "swallowing accelerometry signals" OR "swallowing sounds" OR "swallow sounds" OR "signal processing" OR acoustics OR acoustic OR "doppler effect" OR "doppler shift" OR "sonar doppler" OR microphone)) Tópico:("deglutition" OR "deglutitions" OR "swallowing" OR "swallowings" OR "swallows" OR "swallow") ANDTópico: (videofluoroscopy OR "Videofluoroscopy Swallowing Study" OR VFSS OR "videofluoroscopy study" OR "swallowing videofluoroscopy" OR videofluoroscopic OR "videofluoroscopic swallowing" OR "videofluoroscopic swallowing study" OR fluoroscopy OR fluoroscopies)AND Tópico: (ultrasonography OR ultrasound OR ultrasonics OR "duplex doppler ultrasonography" OR "doppler duplex ultrasonography" OR "acoustic analysis" OR "acoustical analysis" OR auscultation OR "cervical auscultation" OR accelerometry OR "swallowing accelerometry signals" OR "swallowing sounds" OR "swallow sounds" OR "signal processing" OR acoustics OR acoustic OR "doppler effect" OR "doppler shift" OR "sonar doppler" OR microphone)) "deglutition OR swallowing:videofluoroscopy"</p> |
| | | Web of Science | |
| | | Google Scholar | |
| | | Open Grey | Swallowing OR deglutition AND videofluoroscopy AND "acoustic analysis" |
| | | ProQuest | Swallowing OR deglutition AND videofluoroscopy AND "acoustic analysis" |

Appendix 2. Test indicators extracted from De Luca Canto et al.³⁶

| Test indicators | Data analysis | References |
|-----------------|---|---|
| DOR | The value of a DOR ranges from 0 to infinity, with higher values indicating better discriminatory test performance. A value of 1 means that a test does not discriminate between patients with the disorder and those without it. Values lower than 1 point to improper test interpretation (more negative tests among the diseased). | Glas et al. ³⁷ |
| LRs | >1 – diagnostic of interest 0 and 1 – against the diagnosis of interest 0 – less likely the disease =1 – lack diagnostic value | McGee ²⁵ |
| Sensitivity | 80% excellent, 70–80% good, 60–69% fair, <60% poor | No consensus in this regard exists in the literature. |
| Specificity | 90% excellent, 80–90% good, 70–79% fair, <70% poor | No consensus in this regard exists in the literature. |
| Youden's Index | Youden's Index values close to 1 indicate high accuracy; a value of zero is equivalent to uninformed guessing and indicates that a test has no diagnostic value. | Deeks et al. ³⁸ |

Appendix 3. Excluded articles and reasons for exclusion ($n = 22$).

| Author, year | Reason for exclusion |
|--|----------------------|
| Dudik et al., 2016 ³⁹ | 3 |
| Dudik et al., 2015 ⁴⁰ | 3 |
| Dudik et al., 2016 ⁴¹ | 3 |
| Frakking et al., 2016 ⁴² | 3 |
| Frakking et al., 2016 ⁴³ | 3 |
| Golabbakhsh et al., 2014 ⁴⁴ | 3 |
| Lee et al., 2006 ⁴⁵ | 3 |
| Mérey et al., 2012 ⁴⁶ | 3 |
| Morinière et al., 2011 ⁴⁷ | 3 |
| Movahedi et al., 2016 ⁴⁸ | 3 |
| Nikjoo et al., 2011 ³² | 3 |
| Reddy et al., 2000 ⁴⁹ | 3 |
| Sejdic et al., 2014 ⁵⁰ | 3 |
| Seidic et al., 2013 ⁵¹ | 3 |
| Selley et al., 1994 ⁵² | 3 |
| Spadotto et al., 2009 ⁵³ | 3 |
| Spadotto et al., 2008 ⁵⁴ | 3 |
| Steele et al., 2013 ⁵⁵ | 3 |
| Stroud et al., 2002 ⁵⁶ | 3 |
| Tanaka et al., 2012 ⁵⁷ | 3 |
| Zoratto et al., 2010 ⁵⁸ | 3 |
| Lazareck et al., 2004 ⁵⁹ | 3 |

Exclusion criteria: (1) studies in animals; (2) studies that did not performed ultrasound, acoustic analysis, cervical auscultation, swallowing accelerometry signals and doppler effect; (3) studies that do not compare methods of diagnosis of swallowing for both control and dysphagic group with the reference standard (videofluoroscopy); (4) studies that do not present validity measurements (sensitivity and specificity) or did not present data enough to calculate them; (5) reviews, letters, conference, abstract, personal opinions.

Appendix 4. Risk of bias in individual studies. QUADAS-2 criteria fulfilled.

| | Item | Abdulmassih et al., 2013 ¹⁸ | Jayatilake et al., 2015 ¹⁹ | Leslie et al., 2004 ²⁰ |
|------------------------------|---|--|---------------------------------------|-----------------------------------|
| Domain 1: Patient selection | Was a consecutive or random sample of patients enrolled? | N | N | N |
| | Was a case-control design avoided? | N | N | N |
| | Did the study avoid inappropriate exclusions? | Y | Y | Y |
| | Could the selection of patients have introduced bias? | H | H | H |
| | Concerns regarding applicability: Are there concern that the included patients and settings do not match the review question? | L | L | L |
| Domain 2: Index test | Were the index test results interpreted without knowledge of the results of the reference standard? | U | U | U |
| | If a threshold was used, was it prespecified? | Y | Y | Y |
| | Could the conduct or interpretation of the index test have introduced bias? | U | U | L |
| | Concerns regarding applicability: Are there concerns that the index test, its conduct, or its interpretation differ from the review question? | L | L | L |
| Domain 3: Reference standard | Is the reference standard likely to correctly classify the target condition? | Y | Y | Y |
| | Were the reference standard results interpreted without knowledge of the results of the index test? | U | U | U |
| | Could the reference standard, its conduct, or its interpretation have introduced bias? | L | L | L |
| | Concerns regarding applicability: Are there concerns that the target condition as defined by the reference standard does not match the question? | L | L | L |
| Domain 4: Flow and timing | Was there an appropriate interval between index test(s) and reference standard? | Y | Y | Y |
| | Did all patients receive the same reference standard? | Y | Y | Y |
| | Were all patients included in the analysis? | Y | N | N |
| | Could the patient flow have introduced bias? | L | L | L |

Y, yes; N, no; U, unclear; Risk: L, low; H, high; U, unclear.

References

- Logemann JA. Evaluation and treatment of swallowing disorders. *Nat Student Speech Lang Hear Assoc J*. 1984;38–50.
- Miller AJ. The neurobiology of swallowing and dysphagia. *Dev Disabil Res Rev*. 2008;14:77–86.
- Sura L, Madhavan A, Carnaby G, Crary MA. Dysphagia in the elderly: management and nutritional considerations. *Clin Interv Aging*. 2012;7:287–98.
- Chen PH, Golub JS, Hapner ER, Johns MM. Prevalence of perceived dysphagia and quality-of-life impairment in a geriatric population. *Dysphagia*. 2009;24:1–6.
- Smithard DG, O'Neill PA, Parks C, Morris J. Complications and outcome after acute stroke. Does dysphagia matter? *Stroke*. 1996;27:1200–4.
- Roy N, Stemple J, Merrill RM, Thomas L. Dysphagia in the elderly: preliminary evidence of prevalence, risk factors, and socioemotional effects. *Ann Otol Rhinol Laryngol*. 2007;116:858–65.
- Martin-Harris B, Brodsky MB, Michel Y, Castell DO, Schleicher M, Sandidge J, et al. MBS measurement tool for swallow impairment – MBSImp: establishing a standard. *Dysphagia*. 2008;23:392–405.
- Logemann JA. *Manual for the videofluorographic study of swallowing*. 2nd ed. Austin (TX): editor Pro-Ed; 1998.
- Jung SH, Lee LK, Hong JB, Han TR. Validation of clinical dysphagia scale: based on videofluoroscopic swallowing study. *J Korean Acad Rehabil Med*. 2005;29:343–50.
- McCullough GHWR, Rosenbek JC, Mills RH, Webb WG, Ross KB. Inter- and intrajudge reliability for videofluoroscopic swallowing evaluation measures. *Dysphagia*. 2001;16:110–8.
- Scott APA, Bench J. A study of interrater reliability when using videofluoroscopy as an assessment of swallowing. *Dysphagia*. 1998;13:223–7.
- Ian JC, Petter JK. American gastroenterological association technical review on management of oropharyngeal dysphagia. *Gastroenterology*. 1999;455–78.
- Rugiu M. Role of videofluoroscopy in evaluation of neurologic dysphagia. *Acta Otorhinolaryngol Ital*. 2007;27:306–16.
- Bours GJ, Speyer R, Lemmens J, Limburg M, de Wit R. Bedside screening tests vs. videofluoroscopy or fiberoptic endoscopic evaluation of swallowing to detect dysphagia in patients with neurological disorders: systematic review. *J Adv Nurs*. 2009;65:477–93.
- Cichero JA, Murdoch BE. The physiologic cause of swallowing sounds: answers from heart sounds and vocal tract acoustics. *Dysphagia*. 1998;13:39–52.

16. Pehlivan M, Yuceyar N, Ertekin C, Celebi G, Ertaş M, Kalayci T, et al. An electronic device measuring the frequency of spontaneous swallowing: digital phagometer. *Dysphagia*. 1996;11:259–64.
17. Lazareck LJ, Moussavi ZMK. Classification of normal and dysphagic swallows by acoustical means. *IEEE Trans Biomed Eng*. 2004;51:2103–12.
18. David MAL, Alessandro L, Jennifer T, Douglas GA, The PRISMA Group. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *PLoS Med*. 2009;7:1–6.
19. Whiting PF, Rutjes AW, Westwood ME, Mallett S, Deeks JJ, Reitsma JB, et al. QUADAS-2: a revised tool for the quality assessment of diagnostic accuracy studies. *Ann Intern Med*. 2011;155:529–36.
20. Macaskill PGC, Deeks J, Harbord R, Takwoingi Y. Chapter 10: Analysing and presenting results. In: Collaboration C, editor. *Cochrane handbook for systematic reviews of diagnostic test accuracy version 1.0*. England: Birmingham; 2010.
21. Deeks J, Gatsonis C. The Cochrane Collaboration. *Cochrane handbook for systematic reviews of diagnostic test accuracy version 1.0*. Available from: <http://srdta.cochrane.org/>.
22. Abdulmassih EM, Teive HA, Santos RS. The evaluation of swallowing in patients with spinocerebellar ataxia and oropharyngeal dysphagia: a comparison study of videofluoroscopic and sonar Doppler. *Int Arch Otorhinolaryngol*. 2013;17:66–73.
23. Jayatilake D, Ueno T, Teramoto Y, Nakai K, Hidaka K, Ayuzawa S, et al. Smartphone-based real-time assessment of swallowing ability from the swallowing sound. *IEEE J Transl Eng Health Med*. 2015;3:2900310.
24. Leslie P, Drinnan MJ, Finn P, Ford GA, Wilson JA. Reliability and validity of cervical auscultation: a controlled comparison using videofluoroscopy. *Dysphagia*. 2004;19:231–40.
25. McGee S. Simplifying likelihood ratios. *J Gen Intern Med*. 2002;17:647–50.
26. Glas ASLJ, Prins MH, Bonsel GJ, Bossuyt PM. The diagnostic odds ratio: a single indicator of test performance. *J Clin Epidemiol*. 2003;56:1129–35.
27. Palmer JB, Kuhlemeier KV, Tippet DC, Lynch C. A protocol for the videofluorographic swallowing study. *Dysphagia*. 1993;8:209–14.
28. Ramsey DJC, Smithard DG, Kalra L. Early assessments of dysphagia and aspiration risk in acute stroke patients. *Stroke*. 2003;34:1252–7.
29. Dempsey JE, Vice FL, Bosma JF. Combination of cervical auscultation and videoradiography in evaluation of oral and pharyngeal dysphagia. *Abstr Symp Dysphagia*. 1990.
30. Cichero JA, Murdoch BE. Detection of swallowing sounds: methodology revisited. *Dysphagia*. 2002;17:40–9.
31. Sarraf SS, Buchel C, Daun R, Lenton L, Moussavi Z. Detection of swallows with silent aspiration using swallowing and breath sound analysis. *Med Biol Eng Comput*. 2012;50:1261–8.
32. Nikjoo MS, Steele CM, Sejdic E, Chau T. Automatic discrimination between safe and unsafe swallowing using a reputation-based classifier. *Biomed Eng Online*. 2011;10:100.
33. Dudik JM, Kurosu A, Coyle JL, Sejdić E. A comparative analysis of DBSCAN, K-means, and quadratic variation algorithms for automatic identification of swallows from swallowing accelerometry signals. *Comput Biol Med*. 2015;59:10–8.
34. Lagos HN, Santos RS, Abdulmassih EM, Gallinea LF, Langone M. Characterization of swallowing sounds with the use of sonar Doppler in full-term and preterm newborns. *Int Arch Otorhinolaryngol*. 2013;17:383–6.
35. Last JM. A dictionary of epidemiology. 2nd ed. New York: Oxford University Press; 1988.
36. De Luca Canto G, Pachêco-Pereira C, Aydinov S, Major PW, Flores-Mir C, Gozal D. Diagnostic capability of biological markers in assessment of obstructive sleep apnea: a systematic review and meta-analysis. *J Clin Sleep Med*. 2015;11:27–36.
37. Glas AS, Lijmer JG, Prins MH, Bonsel GJ, Bossuyt PM. The diagnostic odds ratio: a single indicator of test performance. *J Clin Epidemiol*. 2003;56:1129–35.
38. Deeks JJ, Bossuyt P, Gatsonis C, editors. *Cochrane handbook for systematic reviews of diagnostic test accuracy version 1.0*. The Cochrane Collaboration. 2010. Available from: <http://srdta.cochrane.org/>.
39. Dudik JM, Coyle JL, Ei-Jaroudi A, Sun MG, Sejdic E. A matched dual-tree wavelet denoising for tri-axial swallowing vibrations. *Biomed Signal Process Control*. 2016;27:112–21.
40. Dudik JM, Kurosu A, Coyle JL, Sejdic E. A comparative analysis of DBSCAN, K-means, and quadratic variation algorithms for automatic identification of swallows from swallowing accelerometry signals. *Comput Biol Med*. 2015;59:10–8.
41. Dudik JM, Kurosu A, Coyle JL, Sejdic E. A statistical analysis of cervical auscultation signals from adults with unsafe airway protection. *J Neuroeng Rehabil*. 2016;13:7.
42. Frakking T, Chang A, O'Grady K, David M, Weir K. Aspiring and nonaspiring swallow sounds in children: a pilot study. *Ann Otol Rhinol Laryngol*. 2016;125:1001–9.
43. Frakking TT, Chang AB, O'Grady KF, David M, Weir KA. Reliability for detecting oropharyngeal aspiration in children using cervical auscultation. *Int J Speech Lang Pathol*. 2016;1–9.
44. Golabbakhsh M, Rajaei A, Derakhshan M, Sadri S, Taheri M, Adibi P. Automated acoustic analysis in detection of spontaneous swallows in Parkinson's disease. *Dysphagia*. 2014;29:572–7.
45. Lee J, Blain S, Casas M, Kenny D, Berall G, Chau T. A radial basis classifier for the automatic detection of aspiration in children with dysphagia. *J Neuroeng Rehabil*. 2006;3:14.
46. Mery C, Kushki A, Sejdic E, Berall G, Chau T. Quantitative classification of pediatric swallowing through accelerometry. *J Neuroeng Rehabil*. 2012;9:34.
47. Moriniere S, Boiron M, Brunereau L, Beutter P, Patat F. Pharyngeal swallowing sound profile assessed after partial and total laryngectomy. *Dysphagia*. 2011;26:366–73.
48. Movahedi F, Kurosu A, Coyle JL, Perera S, Sejdic E. Anatomical directional dissimilarities in tri-axial swallowing accelerometry signals. *IEEE Trans Neural Syst Rehabil Eng*. 2017;25:477–558.
49. Reddy NP, Katakam A, Gupta V, Unnikrishnan R, Narayanan J, Canilang EP. Measurements of acceleration during videofluorographic evaluation of dysphagic patients. *Med Eng Phys*. 2000;22:405–12.
50. Sejdic E, Dudik JM, Kurosu A, Jestrovic I, Coyle JL. Understanding differences between healthy swallows and penetration-aspiration swallows via compressive sensing of tri-axial swallowing accelerometry signals. *Proc SPIE Int Soc Opt Eng*. 2014:9190.
51. Sejdic E, Steele CM, Chau T. Classification of penetration-aspiration versus healthy swallows using dual-axis swallowing accelerometry signals in dysphagic subjects. *IEEE Trans Biomed Eng*. 2013;60:1859–66.
52. Selley WG, Ellis RE, Flack FC, Bayliss CR, Chir B, Pearce VR. The synchronization of respiration and swallow sounds with videofluoroscopy during swallowing. *Dysphagia*. 1994;9:162–7.
53. Spadotto AA, Gatto AR, Guido RC, Montagnoli AN, Cola PC, Pereira JC, et al. Classification of normal swallowing and oropharyngeal dysphagia using wavelet. *Appl Math Comput*. 2009;207:75–82.
54. Spadotto AA, Papa JP, Gatto AR, Cola PC, Pereira JC, Guido RC, et al. Denoising swallowing sound to improve the evaluator's qualitative analysis. *Comput Electr Eng*. 2008;34:148–53.
55. Steele CM, Sejdić E, Chau T. Noninvasive detection of thin-liquid aspiration using dual-axis swallowing accelerometry. *Dysphagia*. 2013;28:105–12.

56. Stroud AE, Lawrie BW, Wiles CM. Inter- and intra-rater reliability of cervical auscultation to detect aspiration in patients with dysphagia. *Clin Rehabil.* 2002;16:640–5.
57. Tanaka N, Nohara K, Okuno K, Kotani Y, Okazaki H, Matsumura M, et al. Development of a swallowing frequency meter using a laryngeal microphone. *J Oral Rehabil.* 2012;39:411–20.
58. Zoratto DCB, Chau T, Steele CM. Hyolaryngeal excursion as the physiological source of swallowing accelerometry signals. *Physiol Meas.* 2010;31:843–55.
59. Lazareck L, Moussavi Z. Classification of normal and dysphagic swallows by acoustical means. *IEEE Trans Biomed Eng.* 2004;51:2103–12.